OSTEOGENICS

REGENERATION PRODUCTS CATALOG

www.osteogenics.com | 888.796.1923

ORDERING

Our customer service professionals are available from 7 AM to 7 PM CST, Monday through Thursday, and 7 AM to 5 PM CST on Fridays. Orders may be placed by the following methods:

TOLL-FREE	1.888.796.1923 (US & Canada only)
INTERNATIONAL	+1 806.796.1923
FAX	806.796.0059
EMAIL	sales@osteogenics.com
WEBSITE	www.osteogenics.com
ADDRESS	Osteogenics Biomedical, Inc. 4620 71st Street | Building 78-79 Lubbock, TX 79424

SHIPPING

Orders placed by 5 PM CST will be shipped the same day unless specified otherwise by your customer service professional. Standard shipping is 2nd Day delivery with UPS. Due to our volume discounts with UPS, our 2nd Day rate is usually less than standard ground shipping and assures better tracking and customer support. Overnight delivery is available at discounted rates as well.

PRICING

Prices are subject to change. However, we will make every effort to notify you in advance of a change. We offer the following discounts on bulk purchases:

Save with our Standard Bulk Discounts!

Buy 5, Get 1 FREE | Buy 12, Get 3 FREE | Buy 30, Get 10 FREE* on all products except Pro-Fix™ kits, Versah®, Meisinger, Master-Mill, and Hu-Friedy products.

**Mixing and matching different products is permitted; the least expensive product will be credited as free.*

PAYMENT

We make it easy for you. We accept all major credit cards, or domestic orders may choose payment terms of Net 15. All payments are in US Dollars.

AVAILABILITY

We know how frustrating back-orders are, so we carry enough inventory to ensure that, statistically, we have your product on hand 99% of the time. In the event of a back-order, we will notify you at the time of your order and give you an estimated ship date.

SATISFACTION ASSURANCE

If you are not completely satisfied with our products, call us and we will arrange for a replacement, exchange, or refund. Unopened boxes may be returned within 30 days from the invoice date for a full refund. Opened boxes may be returned for product exchange within 90 days of the invoice date. Osteogenics also provides a warranty against defects in materials and workmanship on all Pro-Fix™ Precision Fixation System instrumentation for a period of 3 years from date of purchase. Call customer service at 1.888.796.1923 for return authorizations.

OSTEOGENICS

TABLE OF CONTENTS

SOFT TISSUE GRAFTING

4 • Cytoplast MicroDerm™

6 • Zderm™ Collagen Soft Tissue Matrix

BONE GRAFT MATERIALS

8 enCore® Allografts

10 • creos™ allo.gain DBM Putty & DBM Putty with Chips

11 Zcore™ Porcine Xenograft Particulate

12 Zcore™ Form Moldable Collagen-Enriched Porcine Xenograft

13 Zcore™ Expand Collagen-Enriched Porcine Xenograft

14 NovaBone® Dental Putty & Morsels

RESORBABLE BARRIER MEMBRANES

15 Zmatrix™ Porcine Peritoneum Collagen

16 Cytoplast™ RTM Collagen

17 Cytoplast™ RTMPlug, RTMFoam, & RTMTape

NON-RESORBABLE BARRIER MEMBRANES

18 Cytoplast™ Technique Ridge Preservation Kit

18 Cytoplast™ Technique

19 Cytoplast™ TXT-200 & TXT-200 Singles

20 Cytoplast™ Ti-250 & Ti-150 Titanium-Reinforced

24 RPM™ Reinforced PTFE Mesh

SUTURE

26 Cytoplast™ PTFE

27 Resorba® Glycolon™

28 Resorba® Resolon™

28 Resorba® Resolon Twist™

29 Resorba® PGA Resorba™

30 • X-Gut™ Chromic Gut

31 • X-Gut™ Plain Gut

FIXATION SYSTEMS & INSTRUMENTS

32 Pro-Fix™ Membrane Fixation

33 Pro-Fix™ Tenting

34 Pro-Fix™ Bone Fixation

35 Swann-Morton® Premium Micro-Serrated Blades

36 Meisinger

38 • Versah®

BONE SCRAPERS

42 Micross

42 Micross Curve

43 Safescraper® Twist – Curve

43 Safescraper® Twist – Curve Volumizer

44 **SELECTION OF APPLICABLE REFERENCES**

• *New Items Available*

| *All **PART NUMBERS** are denoted with a vertical bar*

NEW

Cytoplast MicroDerm™

Acellular dermal matrix with surface micro-cuts



Dual-sided cross-hatching to increase surface area

Micro-cut channels shorten pathways for vascularization and increase tissue surface area available for attachment.

NOW ENHANCED WITH 6X THE MICRO-CUTS

- Increased surface area at the graft-to-host interface[1]
- Increased cellular infiltration[1]
- Increased hydrophilicity of graft
- Hydrates in 60 seconds
- Ease of trimming and suture placement

TECHNICAL SPECIFICATIONS

- Acellular dermis meets or exceeds all FDA and AATB guidelines for safety
- Terminally sterilized to a SAL of 10^{-6}
- Packaged dehydrated
- No antibiotics, no rinsing
- Not side specific

Case Photos Provided by
Shaun Rotenberg, DMD, MS

1. Pre-op
2. Cytoplast MicroDerm™ placed
3. Sutured
4. Two weeks post-op
5. One year post-op

Site treated with Cytoplast MicroDerm™ via a vestibular tunnel access approach and sutured with Glycolon™.

1. Marinelli et al. Histologic Evaluation of Wound-Bed Preparedness Following Microsurfaced Skin Grafts for the Treatment of Deep Burn Wounds: Results from a Randomized Controlled Trial. Poster presented at 2023 AMSUS Annual Meeting; February 13, 2023; National Harbor, MD.

Cytoplast MicroDerm™

Acellular dermal matrix with surface micro-cuts

NEW

hydrated

"The ***early healing*** and ***soft tissue appearance*** is really ***where Cytoplast MicroDerm™ shines***. Tissue never looks this good so early on with traditional allograft."

Shaun Rotenberg
Periodontist

shown actual size

	Thick (1.75 +/- .25 mm)	Thin (1.25 +/- .25 mm)	
1 cm x 1 cm	CMD1010TK	CMD1010	(1 per box)
1 cm x 2 cm	CMD1020TK	CMD1020	(1 per box)
1 cm x 4 cm	CMD1040TK	CMD1040	(1 per box)
1 cm x 6 cm	CMD1060TK	CMD1060	(1 per box)
2 cm x 4 cm	CMD2040TK	CMD2040	(1 per box)

After hydration, Cytoplast MicroDerm™ becomes soft and pliable.



NEW

Zderm™

Collagen Soft Tissue Matrix

A NEW SOLUTION in SOFT TISSUE GRAFTING DESIGNED TO PERFORM WHEN LEFT EXPOSED*

	ZDERM	LEADING COMPETITOR
Easily handles when dry or hydrated	✓	X
Superior suture retention **strength**[1]	✓	X
Superior tissue **regeneration** at 12 weeks[2]	✓	X
Excellent handling after hydration (product images shown after 30 seconds of hydration)		

1. Suture Retention Strength: Zderm™ (1.67 +/- 0.29) vs. Leading Competitor (1.46 +/- 0.77)
2. Zderm vs. Mucograft Canine Keratinized Tissue Study data on file

PURPOSEFULLY ENGINEERED

Zderm™ Collagen Soft Tissue Matrix is a cross-linked resorbable matrix engineered from highly purified Type I collagen fibers derived from porcine Achilles tendon.

The product is composed of two structures: a coated outer layer that acts as a barrier membrane and a porous matrix layer to allow cell invasion and tissue ingrowth. The product is oriented so that the porous layer is in contact with the host tissue to facilitate tissue integration.

**Primary closure is not required for device performance. Complete wound closure should be attempted when possible.*

INDICATED for:

- *Soft tissue augmentation to increase keratinized tissue*
- *Soft tissue augmentation around implants*



Coated Surface
Magnification x100



Porous Matrix Surface
Magnification x100

Zderm™

Collagen Soft Tissue Matrix

NEW

*The coated front layer should face outward, away from underlying bone. The porous back layer should face inward toward the bone. The front is identified by the location of the chamfer on the **upper left corner** when the Zderm™ is oriented vertically. For the round option, the front layer contains an impression.*



Front View

Coated outer layer that acts as a barrier membrane *- face outward*



Back View

Porous layer allows cell invasion and tissue ingrowth *- face inward*

shown actual size

	Thick (4 - 5 mm)	Thin (1.5 - 3 mm)	
15 mm x 20 mm	ZD1520TK	ZD1520TN	(1 per box)
20 mm x 30 mm	ZD2030TK	ZD2030TN	(1 per box)
30 mm x 40 mm	ZD3040TK	ZD3040TN	(1 per box)
Round 10 mm	ZD10TK		(1 per box)



A SYNERGISTIC COMBINATION

• Combines the synergistic characteristics of slowly resorbing, space-maintaining mineralized cortical bone with osteoinductive demineralized matrix to provide an optimized environment for the regeneration of vital bone

CHAIR-SIDE EFFICIENCY

• 70/30 allograft is pre-mixed to reduce inventory and chair-side preparation
• Double-sterile packaged for aseptic presentation in the surgical field

TESTED TO ENSURE ITS OSTEOINDUCTIVITY

• Pre-sterilization *in vitro* BMP-2 assay
Prior to packaging and terminal sterilization, each lot is tested for a minimum threshold of BMP-2; Lots failing to meet this are discarded

BEST PRACTICES IN SAFETY

• Processed by Allotech, an FDA-registered and AATB accredited tissue bank
• Single donor per lot
• Terminally sterilized by low-dose e-beam irradiation
to a sterility assurance level of 10^{-6}

Representative histology taken at 6 months from a case using combination allograft

86% vital bone
14% residual graft
51% bone, 49% marrow

Histology by Michael Rohrer, DDS, MS
University of Minnesota

enCore® 70|30 Combination Allograft (FDBA & DFDBA)

70% Mineralized Cortical Allograft and 30% Demineralized Allograft

0.25 mm - 1.0 mm Particle Size

0.5 cc	C73050	(1 per box)
1.0 cc	C73100	(1 per box)
1.5 cc	C73150	(1 per box)
2.5 cc	C73250	(1 per box)

enCore® Natural Blend Cortical & Cancellous

Cortical and Cancellous Allograft from a Single Donor

0.25 mm - 1.0 mm Particle Size

0.25 cc	NAT025	(1 per box)
0.5 cc	NAT050	(1 per box)
1.0 cc	NAT100	(1 per box)
1.5 cc	NAT150	(1 per box)
2.5 cc	NAT250	(1 per box)

enCore® 50|50 Cortical & Cancellous Allograft

50% Mineralized Cortical and 50% Mineralized Cancellous

0.5 mm - 1.25 mm Particle Size

0.5 cc	CM55050	(1 per box)
1.0 cc	CM55100	(1 per box)
1.5 cc	CM55150	(1 per box)
2.5 cc	CM55250	(1 per box)

enCore® OD 30|70 Cortical & Cancellous Allograft

30% Mineralized Cortical Allograft and 70% Mineralized Cancellous Allograft

0.25 mm - 0.85 mm Particle Size

0.5 cc	OD37050	(1 per box)
1.0 cc	OD37100	(1 per box)
2.5 cc	OD37250	(1 per box)

enCore® Cortical Allograft

0.25 mm - 1.0 mm Particle Size

0.25 cc	SMIN025	(1 per box)
0.5 cc	SMIN050	(1 per box)
1.0 cc	SMIN100	(1 per box)
1.5 cc	SMIN150	(1 per box)
2.5 cc	SMIN250	(1 per box)

enCore® Cancellous Allograft

0.25 mm - 1.0 mm Particle Size

0.25 cc	MCAN025	(1 per box)
0.5 cc	MCAN050	(1 per box)
1.0 cc	MCAN100	(1 per box)
1.5 cc	MCAN150	(1 per box)
2.5 cc	MCAN250	(1 per box)

enCore® Demineralized Cortical Allograft

0.125 mm - 0.85 mm Particle Size

0.25 cc	DMCR025	(1 per box)
0.5 cc	DMCR050	(1 per box)
1.0 cc	DMCR100	(1 per box)
1.5 cc	DMCR150	(1 per box)
2.5 cc	DMCR250	(1 per box)

NEW

creos™ allo.gain

DBM Putty & DBM Putty with Chips

creos™ allo.gain DBM (demineralized bone matrix) putty is a 100% pure demineralized allograft. It resists irrigation and comes in an easy-application syringe. creos™ allo.gain DBM putty with chips is a cohesive, moldable bone graft composed of demineralized human cortical bone matrix and cancellous chips. Both products retain native growth factors inherent to bone tissue and are provided sterile and ready for implantation.

CARRIER-FREE SOLUTION

creos™ allo.gain DBM putty is carrier-free and intended for use in dental surgeries as an osteoinductive and osteoconductive bone void filler. Including:

- Bone defect filling
- Dental socket preservation
- Ridge augmentation

VERIFIED OSTEOINDUCTIVE POTENTIAL

Every lot of allo.gain™ DBM putty is verified through *in vitro* testing to meet a statistically validated BMP threshold for osteoinductivity, providing clinicians with confidence in its ability to support new bone formation.

NO ADDED FILLERS

This product contains only human DBM, human cancellous chips, and no additional processing agents remain in the final crush-mix.

DBM Putty

0.5 cc	**N6110A**	(1 per box)
1.0 cc	**N6120A**	(1 per box)
2.5 cc	**N6130A**	(1 per box)

DBM Putty with Chips

1.0 cc	**N6220**	(1 per box)
2.5 cc	**N6230**	(1 per box)

not actual size

Zcore™

Porcine Xenograft Particulate

Zcore™ is an osteoconductive, porous, anorganic bone mineral with a carbonate apatite structure derived from porcine cancellous bone, eliminating risk of BSE transmission.

Zcore™ Porcine Xenograft Particulate

.25 mm - 1.0 mm Particle Size

0.5 cc	**ZS050**	(1 per box)
1.0 cc	**ZS100**	(1 per box)
2.0 cc	**ZS200**	(1 per box)
4.0 cc	**ZS400**	(1 per box)

1.0 mm - 2.0 mm Particle Size

1.0 cc	**ZL100**	(1 per box)
2.0 cc	**ZL200**	(1 per box)

not actual size

Zcore™ Porcine Xenograft Particulate in Syringe

.25 mm - 1.0 mm Particle Size

0.25 cc	**ZY025**	(1 per box)
0.5 cc	**ZY050**	(1 per box)



SEM of Processed Human Bone
Magnification x50



SEM of Zcore™ Porcine Xenograft Particulate
Magnification x50

INTERCONNECTING PORES

Interconnecting macroscopic and microscopic porous structure supports the formation and ingrowth of new bone

88% TO 95% VOID SPACE

88% to 95% Void Space: hyper-porosity of porcine cancellous matrix and intra-particle space facilitated by rough particle morphology reduce bulk density of the graft, allowing greater empty space for new bone growth*

**0.25 mm - 1.0 mm particle size = 88% void space, 1.0 mm - 2.0 mm = 95% void space*

PROCESSED USING MINIMAL HEAT

Heat treated to an optimal temperature that ensures a degree of crystallinity[1] consistent with native bone mineral to allow for remodeling of the healing bone

1. Li ST, Chen HC, Yuen D. Isolation and Characterization of a Porous Carbonate Apatite From Porcine Cancellous Bone. Science, Technology, Innovation, Aug. 2014: 1-13.

Zcore™ Form

Moldable Collagen-Enriched Porcine Xenograft

shown actual size

9 mm diam. x 8 mm

0.5 cc | **ZF050** (1 per box)

11 mm diam. x 12 mm

1.0 cc | **ZF100** (1 per box)

11 mm diam. x 22 mm

2.0 cc | **ZF200** (1 per box)

Zcore™ Form hydrates almost immediately when introduced to the patient's blood or sterile saline.

Once hydrated, Zcore™ Form becomes moldable and can take the shape of a variety of defect shapes and sizes.

90% ZCORE™ PORCINE XENOGRAFT PARTICULATE
10% TYPE I PORCINE COLLAGEN

A composite of osteoconductive bone mineral and collagen, Zcore™ Form is composed of 90% porcine xenograft particulate and 10% porcine collagen by weight. The moldable consistency allows it to take the shape of the defect while also making the overall handling of the product easier and more convenient than particulate grafts.



Zcore™ Expand

Expandable Collagen-Enriched Porcine Xenograft

Zcore™ Expand hydrates and expands almost immediately when introduced to the patient's blood or sterile saline.

Once hydrated, Zcore™ Expand increases in diameter to fill the void space in a socket or sinus defect.





shown actual size

Socket

5 mm x 17 mm

10 mm x 17 mm EXPANDED

ZXSOCKET	(1 unit)
ZXSOCKET-5	(5 units)
ZXSOCKET-10	(10 units)
ZXSOCKET-20	(20 units)



Small Sinus

13 mm diam. x 10 mm

17 mm x 10 mm EXPANDED

ZXSINUSS	(1 unit)



Large Sinus

17 mm diam. x 12 mm

22 mm x 12 mm EXPANDED

ZXSINUSL	(1 unit)

80% ZCORE™ PORCINE XENOGRAFT PARTICULATE
20% TYPE I EXPANDABLE PORCINE COLLAGEN

A composite of osteoconductive bone mineral and expandable collagen, Zcore™ Expand is composed of 80% porcine xenograft particulate and 20% porcine collagen by weight. Zcore™ Expand is supplied as a compressed preformed sponge that expands when hydrated, allowing it to take the shape of the defect. The unique expandable property makes Zcore™ Expand a desirable option for extraction site grafting and/or sinus augmentation that uses a lateral approach.

NovaBone® Dental Putty & NovaBone® Morsels

The synthetic solution to bone regeneration

NovaBone® Putty in Cartridges

Cartridges

0.25 cc	NA4640	(4 per box)
0.5 cc	NA3620	(2 per box)
0.5 cc	NA3660	(6 per box)



Cartridge Applicator Gun

NA4600	(Fits all cartridges)

NovaBone® Putty in Syringes

0.5 cc	NA1610	(1 per box)
1.0 cc	NA1611	(1 per box)
2.0 cc	NA1612	(1 per box)

"It's ***amazing*** for vertical approach sinus lifts in conjunction with implant placement! It lifts the membrane more predictability than other graft materials I've used, and it's more apparent on the x-ray due to radiopacity. I'm very happy and impressed with NovaBone™! I now feel I have greater predictability with vertical approach sinus lifts, and I'm doing it in situations when I would have previously used a lateral window approach to the sinus lift. ***The results have been fantastic!***"

Scott Price, DDS
Periodontist

NovaBone® Morsels in Trays

1.3 cc	EU0820	(2 per box)

not actual size

NovaBone® Morsels is a particulate product made up of a crystalline composite calcium phosphosilicate (CPS). The particle size ranges from 0.5 mm - 1.0 mm with pore sizes ranging from 0.05 mm - 0.10 mm. The pore size results in slow and sustained resorption that is completed over a 12-18 month period. The morsels have an "osteostimulative" effect similar to NovaBone® Dental Putty.

Zmatrix™

Porcine peritoneum collagen membrane

A perfectly soft consistency that drapes without the usual self-adherence experienced with other natural collagen membranes.

shown actual size

15 mm x 20 mm

| ZM1520 (1 per box)

25 mm x 30 mm

| ZM2530 (1 per box)

30 mm x 40 mm

| ZM3040 (1 per box)

EASY HANDLING

- Does not stick to instruments or itself
- Either side can face the defect
- Low surface expansion when hydrated

HIGH TENSILE STRENGTH[1]

- High suture retention
- High tear resistance

FACILITATES NEW BONE FORMATION[2, 3]

- Significantly higher new bone formation in the central portion of the defect, in comparative *in vivo* study

1. Gasser A, et al. J Dent Res 2016,95 (Spec Iss A): 1683 2. Wessing B, et al. Clin Oral Impl Res; 2017;28(11):e218-e226 3. Omar O, et al. Clin Oral Impl Res; 2018;29(1):7-19

Cytoplast™ RTM Collagen

Type I bovine collagen membrane

shown actual size

15 mm x 20 mm

| RTM1520 (2 per box)

20 mm x 30 mm

| RTM2030 (2 per box)

30 mm x 40 mm

| RTM3040 (2 per box)

"...I am impressed with its ***handling***, but most importantly, I am impressed with its ***results***."

Jerald Rosenberg, DMD
Periodontist

MANUFACTURED FROM HIGHLY PURIFIED TYPE 1 BOVINE ACHILLES TENDON

Safe for the patient

26 – 38 WEEK RESORPTION TIME

Long predictable resorption time limits the risk of particle loss due to premature resorption

HIGH TENSILE STRENGTH

You can suture or tack the membrane in place without tearing

CELL OCCLUSIVE

Prevents epithelial down growth

OPTIMIZED FLEXIBILITY

Stiff enough for easy placement, yet easily drapes over ridge



Reconstituted fiber construction allows tissue integration while preventing direct passage of epithelial cells.

Cytoplast™ RTMPlug, RTMFoam, & RTMTape

Absorbable Wound Dressing | Type I & Type III bovine collagen

shown actual size

RTMPlug

1 cm x 2 cm

| RTMPLUG10 (10 per box)

RTMFoam

2 cm x 4 cm (3 mm thick)

| RTMFOAM10 (10 per box)

RTMTape

2.5 cm x 7.5 cm (1 mm thick)

| RTMTAPE10 (10 per box)

Wound dressings will be essentially resorbed within 30 days

APPLICATIONS

- Surgical wounds
- Periodontal surgical wounds
- Extraction sites
- Dental sores
- Oral ulcers (non-infected or viral)
- Suture sites
- Burns
- Traumatic wounds

Cytoplast™ Technique

Ridge preservation without primary closure | U.S. Patent # 6,019,764

Cytoplast™ Technique Ridge Preservation Kit

| KITRPCT

1 **enCore® 70/30 Combination Allograft** *0.5 cc*

1 **Cytoplast™ TXT-200 Single**

1 **Cytoplast™ PTFE Suture** *USP 3/0; 16 mm RC needle*



1. Preoperative view. To maximize the result of ridge preservation procedures, techniques designed to minimize trauma to the alveolar bone, such as the use of periotomes and surgical sectioning of ankylosed roots should be considered.

2. All soft tissue remnants should be removed with sharp curettage. Special care should be taken to remove all soft tissue at the apical extent of the socket of endodontically treated teeth. Bleeding points should be noted on the cortical plate. If necessary, decortication of the socket wall should be done with a #2 round burr to improve blood supply.

3. A subperiosteal pocket is created with a micro periosteal elevator or small curette, extending 3-5 mm beyond the socket margins on the palatal and the facial aspect of the socket. In the esthetic zone, rather than incising and elevating the interdental papilla, it is left intact and undermined in a similar fashion. The Cytoplast™ high-density PTFE membrane will be tucked into this subperiosteal pocket.

4. Particulate graft material can be placed into the socket with a syringe or with a curette. Ensure that the material is evenly distributed throughout the socket. However, the particles should not be densely packed to preserve ample space for blood vessel ingrowth.

5. The Cytoplast™ high-density PTFE membrane is trimmed to extend 3-5 mm beyond the socket walls and then tucked subperiosteally under the palatal flap, the facial flap and underneath the interdental papilla with a curette. The membrane should rest on bone 360° around the socket margins, if possible. Note that minimal flap reflection is necessary to stabilize the membrane.

6. Ensure that there are no folds or wrinkles in the membrane and that it lies passively over the socket. To prevent bacterial leakage under the membrane, take care to avoid puncturing the membrane, and do not overlap two adjacent pieces of membrane material.

7. The membrane is further stabilized with a criss-cross Cytoplast™ PTFE suture. Alternatively, interrupted sutures may be placed. The PTFE sutures, which cause minimal inflammatory response, are left in place for 10 to 14 days.

8. The membrane is removed, non-surgically, in 21 to 28 days. Sockets with missing walls may benefit from the longer time frame. Topical anesthetic is applied, then the membrane is grasped with a tissue forcep and removed with a gentle tug.

9. Studies have shown that by 21-28 days there is a dense, vascular connective tissue matrix in the socket and early osteogenesis is observed in the apical 2/3 of the socket.

10. Immediately following membrane removal, a dense, highly vascular, osteoid matrix is observed. The natural position of the gingival margin has been left intact because primary closure was not necessary. The dense PTFE membrane has contained the graft material and prevented epithelial migration into the socket.

11. The socket at 6 weeks. Keratinized gingiva is beginning to form over the grafted socket. The natural soft tissue architecture is preserved, including the interdental papillae. New bone is beginning to form in the socket. Over the next 6 to 10 weeks, increasing thickness of trabeculae and mineralization will result in load bearing bone suitable for implant placement.

Cytoplast™ TXT-200 & TXT-200 Singles

Micro-textured, high-density PTFE membrane

"I always know, ***in advance***, the results of my bone grafting when I use Cytoplast™ TXT-200 as a membrane. ***Why bother with other membranes?***"

Mark Cohen, DDS
Periodontist

shown actual size

MOST POPULAR SOCKET GRAFTING MEMBRANE

TXT-200 Singles

12 mm x 24 mm

- **TXT1224-1** (1 per box)
- **TXT1224** (10 per box)

TXT-200

12 mm x 30 mm

- **TXT1230** (10 per box)

TXT-200

25 mm x 30 mm

- **TXT2530-1** (1 per box)
- **TXT2530** (4 per box)



The patented Regentex™ surface helps stabilize the membrane and the soft tissue flap. Hexagonal surface dimples provide a textured surface that increases the area available for cellular attachment without increasing porosity. U.S. Patent #5,957,690

NON-RESORBABLE

Won't resorb prematurely – you dictate healing time

100% DENSE (NON-EXPANDED) PTFE

Impervious to bacteria (pore size less than 0.3 µm) *Data on file*

PURPOSELY LEAVE THE MEMBRANE EXPOSED

Preservation of the soft tissue architecture and keratinized mucosa

SOFT TISSUE ATTACHES, BUT DOESN'T GROW THROUGH THE MEMBRANE

Exposed membrane allows for non-surgical removal; no anesthesia required

HEXAGONAL DIMPLES INCREASE SURFACE AREA

Designed to increase membrane stabilization

Cytoplast™ Titanium-Reinforced

Titanium-reinforced, high-density PTFE membrane

Ti-250 (250 µm thick)	Ti-150 (150 µm thick)	

ANL

12 mm x 24 mm

Ti-250 (250 µm thick)	Ti-150 (150 µm thick)	
Ti250ANL-N-1	Ti150ANL-N-1	(1 per box)
Ti250ANL-N-2	Ti150ANL-N-2	(2 per box)

Designed for narrow single-tooth extraction sites, especially where one bony wall is missing

ANL30

12 mm x 30 mm

Ti250ANL30-N-1		(1 per box)
Ti250ANL30-N-2		(2 per box)

Designed for narrow single-tooth extraction sites, especially where one bony wall is missing

PS

20 mm x 25 mm

Ti250PS-N-1	Ti150PS-N-1	(1 per box)
Ti250PS-N-2	Ti150PS-N-2	(2 per box)

Designed for large extraction sites and limited ridge augmentation

PL

25 mm x 30 mm

Ti250PL-N-1	Ti150PL-N-1	(1 per box)
Ti250PL-N-2	Ti150PL-N-2	(2 per box)

Designed for large bony defects, including ridge augmentation

VERSATILE RECTANGULAR SHAPES

These configurations can be trimmed to fit a variety of defects.
Shown actual size.

Cytoplast™ Titanium-Reinforced

Titanium-reinforced, high-density PTFE membrane

**Ti-150 membranes are 40% thinner than Ti-250 membranes, providing clinicians another handling option in Cytoplast™ Titanium-Reinforced Membranes.*

VERSATILE RECTANGULAR SHAPES
These configurations can be trimmed to fit a variety of defects.
Shown actual size.

Ti-250 (250 µm thick)	Ti-150 (150 µm thick)	

XL

30 mm x 40 mm

Ti-250 (250 µm thick)	Ti-150 (150 µm thick)	
Ti250XL-N-1	Ti150XL-N-1	(1 per box)
Ti250XL-N-2	Ti150XL-N-2	(2 per box)

Designed for very large bony defects, including ridge augmentation

XLK

30 mm x 40 mm

Ti-250 (250 µm thick)	Ti-150 (150 µm thick)	
Ti250XLK-N-1	Ti150XLK-N-1	(1 per box)
Ti250XLK-N-2	Ti150XLK-N-2	(2 per box)

Designed for very large bony defects, including ridge augmentation

K2

40 mm x 50 mm

Ti-250 (250 µm thick)	Ti-150 (150 µm thick)	
Ti250K2-N-1	Ti150K2-N-1	(1 per box)
Ti250K2-N-2	Ti150K2-N-2	(2 per box)

Designed for the largest bony defects, including ridge augmentation

Cytoplast™ Titanium-Reinforced

Titanium-reinforced, high-density PTFE membrane

Ti-250 (250 µm thick)	Ti-150 (150 µm thick)	
AS		
14 mm x 24 mm		
Ti250AS-N-1	Ti150AS-N-1	(1 per box)
Ti250AS-N-2	Ti150AS-N-2	(2 per box)

Designed for single-tooth extraction sites, especially where one or more bony walls are missing

ATC		
24 mm x 38 mm		
Ti250ATC-N-1		(1 per box)
Ti250ATC-N-2		(2 per box)

Designed for large extraction sites, including ridge augmentation

Dimensional measurements shown in mm. Width measurements noted at widest point and narrowest point. Shown actual size.

INTERPROXIMAL SHAPES

These configurations are designed to fit between existing teeth.

Cytoplast™ Titanium-Reinforced

Titanium-reinforced, high-density PTFE membrane

Dimensional measurements shown in mm. Width measurements noted at widest point and narrowest point. Shown actual size.

SHAPES WITH FIXATION POINTS

These configurations are designed with fixation points outside of the defect area.







Ti-250 (250 µm thick)	Ti-150 (150 µm thick)	

BL

17 mm x 25 mm

Ti250BL-N-1	Ti150BL-N-1	(1 per box)
Ti250BL-N-2	Ti150BL-N-2	(2 per box)

Designed for large buccal defects

NEW BLL

17 mm x 30 mm

Ti250BLL-N-1	Ti150BLL-N-1	(1 per box)
Ti250BLL-N-2	Ti150BLL-N-2	(2 per box)

Designed for large buccal defects

PST

36 mm x 25 mm

Ti250PST-N-1	Ti150PST-N-1	(1 per box)
Ti250PST-N-2	Ti150PST-N-2	(2 per box)

Designed for large extraction sites and limited ridge augmentation in the anterior maxilla

PLT

41 mm x 30 mm

Ti250PLT-N-1	Ti150PLT-N-1	(1 per box)
Ti250PLT-N-2	Ti150PLT-N-2	(2 per box)

Designed for large bony defects, including ridge augmentation in the anterior maxilla

RPM™

Reinforced PTFE mesh

CIRCULAR MACROPORES

Allow direct contact between the bone graft and periosteum, allowing naturally occurring revascularization and infiltration of cells into the bone graft

TITANIUM FRAME

Maintains space essential for horizontal and vertical ridge augmentation

PTFE MESH

Easily conforms to tissue contours



Hybrid Approach: *Adaptability of a membrane with the porosity of a mesh*

SHAPES WITH FIXATION POINTS

These configurations are designed with fixation points outside of the defect area.



BL

17 mm x 25 mm

| RPM200BL (1 per box)

Designed for large buccal defects

BLL

17 mm x 30 mm

| RPM200BLL (1 per box)

Designed for large buccal defects

PST

36 mm x 25 mm

| RPM200PST (1 per box)

Designed for large extraction sites and limited ridge augmentation in the anterior maxilla

PLT

41 mm x 30 mm

| RPM200PLT (1 per box)

Designed for large bony defects, including ridge augmentation in the anterior maxilla

RPM™

Reinforced PTFE mesh

VERSATILE SHAPES

These configurations can be trimmed to fit a variety of defects. Shown actual size.

INTERPROXIMAL

This configuration is designed to fit between existing teeth.

PS

20 mm x 25 mm

| RPM200PS (1 per box)

Designed for large extraction sites and limited ridge augmentation

PL

25 mm x 30 mm

| RPM200PL (1 per box)

Designed for large bony defects, including ridge augmentation

XLK

30 mm x 40 mm

| RPM200XLK (1 per box)

Designed for very large bony defects, including ridge augmentation

XL

30 mm x 40 mm

| RPM200XL (1 per box)

Designed for very large bony defects, including ridge augmentation

K2

40 mm x 50 mm

| RPM200K2 (1 per box)

Designed for the largest bony defects, including ridge augmentation

ATC

24 mm x 38 mm

| RPM200ATC (1 per box)

Designed for large extraction sites, including ridge augmentation

Cytoplast™ PTFE Suture

The soft monofilament suture

All Cytoplast™ Sutures are 12 per box
Available in 18" and 28" lengths

18" Undyed 28" Undyed	**Precision RC** 19 mm	2/0 USP	**CS0418** **CS0428**
18" Undyed 28" Undyed	**Precision RC** 16 mm	3/0 USP	**CS0518** **CS0528**
18" Undyed 28" Undyed	**Precision RC** 19 mm	3/0 USP	**CS051819** **CS052819**
18" Undyed 28" Undyed	**TP** 13 mm	4/0 USP	**CS0618PERIO** **CS0628PERIO**
18" Undyed 28" Undyed	**Precision RC** 13 mm	4/0 USP	**CS0618PREM** **CS0628PREM**
18" Undyed 28" Undyed	**Precision RC** 16 mm	4/0 USP	**CS0618RC** **CS0628RC**
18" Undyed 28" Undyed	**Precision RC** 13 mm	5/0 USP	**CS071813** **CS072813**
18" Undyed 28" Undyed	**Precision RC** 16 mm	5/0 USP	**CS071816** **CS072816**

CYTOPLAST™ NEEDLE CODE DETAIL

RC 3/8 CIRCLE REVERSE CUTTING

TP 1/2 CIRCLE ROUND-BODIED

100% MEDICAL GRADE PTFE
Biologically inert

MONOFILAMENT
Doesn't wick bacteria

SOFT (NOT STIFF)
Comfortable for patients

LITTLE TO NO PACKAGE MEMORY
Excellent handling, knots securely

NON-RESORBABLE
Keeps the surgical site reliably closed

300 SERIES STAINLESS STEEL NEEDLES

All Cytoplast™ PTFE Sutures have 300 series stainless steel needles, the gold standard material for suture needles. Tests comparing the 300 series needles to previous needles show a substantial increase in needle strength, initial needle sharpness, and sustained needle sharpness. Tests show that they are less likely to bend, require less force to penetrate, and maintain sharpness longer. Additionally, all silver needles now have longer and geometrically finer precision cutting edges. *Data on file*

Resorba® Glycolon™

Absorbable, Monofilament

Glycolon™ is Resorba's® top selling suture material worldwide and is comprised of polyglycolic acid (PGA) and polycaprolactone (PCL). The monofilament structure provides excellent handling properties, does not wick bacteria, and allows for atraumatic passage through the tissue. Glycolon™ maintains 50% of its tensile strength for 11-13 days. *In Vivo data on file*

Violet	**DSM** 16 mm	3/0 USP	**OD01204**
Violet	**DSM** 18 mm	3/0 USP	**OD01205**
Violet	**HRT** 18 mm	4/0 USP	**OD01101**
Violet	**DSM** 16 mm	4/0 USP	**OD01201**
Violet	**DSM** 18 mm	4/0 USP	**OD01203**
Violet	**DSM** 13 mm black needle	5/0 USP	**OD01210**
Violet	**DSM** 16 mm black needle	5/0 USP	**OD01211**
Violet	**DSM** 18 mm black needle	5/0 USP	**OD01212**
Violet	**HRT** 16 mm	5/0 USP	**OD01100**
Violet	**DSM** 16 mm	5/0 USP	**OD01214**
Undyed	**DSM** 18 mm	5/0 USP	**OD01202**
Undyed	**DSM** 13 mm	6/0 USP	**OD01200**
Violet	**DSM** 13 mm	6/0 USP	**OD01213**

MICRO SUTURE:

Violet	**HRT** 10 mm	6/0 USP	**OD01102**

"I like it so much that if I ask for a suture, ***my staff doesn't need to ask what I want...they know it's 6-0 Glycolon™.*** Can't beat that."

Israel Puterman, DMD, MSD
Periodontist

"***Hands down my favorite resorbable sutures***; very easy to handle, ***so clean on post ops***, also I love that it stays for a long time with good tensile strength. Last, but not least, it looks so beautiful in pictures."

Thaer Alqadoumi, DDS
Periodontist

RESORBA™ NEEDLE CODE DETAIL

DSM 3/8 CIRCLE PREMIUM REVERSE CUTTING
DS 3/8 CIRCLE STANDARD REVERSE CUTTING
HRT 1/2 CIRCLE ROUND-BODIED CUTTING
HR 1/2 CIRCLE ROUND-BODIED
HSM 1/2 CIRCLE PREMIUM REVERSE CUTTING
HS 1/2 CIRCLE STANDARD REVERSE CUTTING



Resorba® Resolon™

Non-Absorbable, Monofilament

Resolon™ is initially like traditional nylon sutures until it undergoes a proprietary treatment process that results in a softer and more supple version of a nylon suture. Resolon™ provides clinicians a non-absorbable monofilament suture option that does not wick bacteria and has superior handling characteristics when compared to traditional nylon sutures.

Blue	**DSM** 13 mm	4/0 USP	**OD13202**
Blue	**DSM** 16 mm	4/0 USP	**OD13205**
Blue	**DSM** 18 mm	4/0 USP	**OD13207**
Blue	**DSM** 16 mm black needle	4/0 USP	**OD13215**
Blue	**HS** 18 mm	5/0 USP	**OD13700**
Blue	**DSM** 13 mm	5/0 USP	**OD13201**
Blue	**DSM** 16 mm	5/0 USP	**OD13204**
Blue	**DSM** 18 mm	5/0 USP	**OD13206**
Blue	**DSM** 13 mm black needle	5/0 USP	**OD13213**
Blue	**DSM** 16 mm black needle	5/0 USP	**OD13214**
Blue	**DSM** 13 mm	6/0 USP	**OD13200**
Blue	**DSM** 16 mm	6/0 USP	**OD13203**
Blue	**DSM** 13 mm black needle	6/0 USP	**OD13212**

MICRO SUTURE:

Blue	**DSM** 13 mm black needle	7/0 USP	**OD13211**

Resorba® Resolon Twist™

Non-Absorbable, Pseudo-Monofilament

Resolon Twist™ is a pseudo-monofilament made of braided nylon fibers that are coated with a nylon sheath. The pseudo-monofilament design offers clinicians a non-absorbable suture that handles similarly to a multifilament suture but, due to its outer nylon coating, has the advantage of reduced drag when being pulled through soft tissue.

Undyed	**HRT** 18 mm	3/0 USP	**OD12100**
Undyed	**DSM** 18 mm	4/0 USP	**OD12201**
Undyed	**DSM** 16 mm	4/0 USP	**OD12200**
Undyed	**HS** 15 mm	4/0 USP	**OD12700**

Resorba® PGA Resorba™

Absorbable, Multifilament

PGA Resorba™ is an absorbable suture made of precision-braided filaments of polyglycolic acid coated with a special resolactone coating to reduce surface friction when passing through tissue. The composition of PGA Resorba™ ensures predictable and moderately rapid resorption in tissue. PGA Resorba™ maintains 50% tensile strength for up to 21 days. *In Vivo data on file*

Violet	**HRT** 18 mm	4/0 USP	**OD03100**
Violet	**DSM** 18 mm	4/0 USP	**OD03202**
Violet	**HR** 17 mm	5/0 USP	**OD03500**
Violet	**DS** 18 mm	5/0 USP	**OD03400**
Violet	**DSM** 13 mm	5/0 USP	**OD03201**
Violet	**DSM** 13 mm	6/0 USP	**OD03200**

MICRO SUTURE:

Violet	**HRT** 10 mm	5/0 USP	**OD03103**
Violet	**HSM** 10 mm	5/0 USP	**OD03700**
Violet	**HSM** 10 mm	6/0 USP	**OD03701**
Violet	**DSM** 11 mm	6/0 USP	**OD03203**
Violet	**DSM** 7 mm	7/0 USP	**OD03206**
Violet	**DSM** 11 mm	7/0 USP	**OD03204**



NEW

X-Gut™ Chromic Gut Suture

Maintains tensile strength for 14 – 21 days

All X-Gut™ Sutures are 12 per box

All needles are 3/8 Circle except where indicated

45 cm Chromic Gut	**RC** 19 mm	USP 3/0	**GC301945**
75 cm Chromic Gut	**RC** 19 mm	USP 3/0	**GC301975**
45 cm Chromic Gut	**Precision RC** 19 mm	USP 3/0	**GC3019P45**
75 cm Chromic Gut	**Precision RC** 19 mm	USP 3/0	**GC3019P75**
45 cm Chromic Gut	**RC** 19 mm	USP 4/0	**GC401945**
75 cm Chromic Gut	**RC** 19 mm	USP 4/0	**GC401975**
45 cm Chromic Gut	**Precision RC** 13 mm	USP 4/0	**GC4013P45**
45 cm Chromic Gut	**Precision RC** 19 mm	USP 4/0	**GC4019P45**
45 cm Chromic Gut	**Precision RC** 13 mm	USP 5/0	**GC5013P45**
45 cm Chromic Gut	**RC** 19 mm	USP 5/0	**GC501945**
45 cm Chromic Gut	**Precision RC** 16 mm	USP 5/0	**GC5016P45**

X-Gut™ Chromic Gut suture is an absorbable suture composed of purified connective tissue (mostly collagen) derived from the serosal layer of beef (bovine) intestines.

X-Gut™ Chromic Gut sutures maintain tensile strength for 14 – 21 days.

X-Gut™ Chromic Gut sutures have a high tensile strength compared to other gut sutures, minimizing the risk of tearing during suturing.

X-Gut™ Plain Gut Suture

Maintains tensile strength for about 10 days

NEW

X-Gut™ Plain Gut suture is an absorbable suture composed of purified connective tissue (mostly collagen) derived from the serosal layer of beef (bovine) intestines.

X-Gut™ Plain Gut sutures maintain tensile strength for about 10 days.

All X-Gut™ Sutures are 12 per box

All needles are 3/8 Circle except where indicated

45 cm Plain Gut	**RC** 19 mm	USP 3/0	**| GP301945**
75 cm Plain Gut	**RC** 19 mm	USP 3/0	**| GP301975**
45 cm Plain Gut	**1/2 Circle RC** 22 mm	USP 3/0	**| GP302245**
75 cm Plain Gut	**RC** 24 mm	USP 3/0	**| GP302475**
45 cm Plain Gut	**RC** 19 mm	USP 4/0	**| GP401945**
75 cm Plain Gut	**RC** 19 mm	USP 4/0	**| GP401975**

X-GUT™ CHROMIC & PLAIN NEEDLE CODE DETAIL

RC | PRECISION RC	3/8 CIRCLE REVERSE CUTTING
1/2 CIRCLE RC	1/2 CIRCLE REVERSE CUTTING

Pro-Fix™ Membrane Fixation

Precision Fixation System

Tray and organizer dial are designed to store all Pro-fix™ components including up to 100 membrane fixation, tenting, and bone fixation screws. Blades are designed to work universally with all Pro-fix™ membrane fixation, tenting, and bone fixation screws.

Pro-fix™ Membrane Fixation Screws are designed as an attractive alternative to using tacks for membrane stabilization. Easy pick-up, solid stability of the screw during transfer to the surgical site, and easy placement make membrane fixation fast and easy.

Membrane Fixation Kit | PFMK20

- Autoclavable Tecapro™ storage tray w/ screw organizer dial
- Stainless steel driver handle
- 76 mm cruciform driver blade
- 56 mm cruciform driver blade
- (20) 1.5 x 3 mm self-drilling membrane fixation screws

Individual Components

Stainless Steel Driver Handle	PFDH
76 mm Cruciform Driver Blade	PFDB
56 mm Cruciform Driver Blade	PFDB56
24 mm Contra Angle Blade (10 mm exposed distal length)	PFDBCA
1.2 mm diam. Latch Type Pilot Drill	BI1001
Autoclavable Tecapro™ storage tray	PFT

Self-Drilling Membrane Fixation Screws

1.5 mm x 3 mm

PFMF-5	(5 per box)
PFMF-10	(10 per box)
PFMF-20	(20 per box)

actual size

1.5 mm x 5 mm

PFMF5-5	(5 per box)
PFMF5-10	(10 per box)
PFMF5-20	(20 per box)

actual size

Pro-Fix™ Tenting

Precision Fixation System

Pro-fix™ Tenting Screws are designed with a self-drilling tip, polished neck, and broader head to maintain space under resorbable and non-resorbable membranes in horizontal and vertical bone regeneration procedures.

Tenting Kit

| PFTK12

- Autoclavable Tecapro™ storage tray w/ screw organizer dial
- Stainless steel driver handle
- 76 mm cruciform driver blade
- 56 mm cruciform driver blade
- (4) **1.5 x 3 mm self-drilling tenting screws** (7 mm total length)
- (4) **1.5 x 4 mm self-drilling tenting screws** (8 mm total length)
- (4) **1.5 x 5 mm self-drilling tenting screws** (9 mm total length)

For individual Pro-Fix™ driver and container components, see opposite page.

Self-Drilling Tenting Screws

actual size

1.5 mm x 3 mm *polished neck*

+ 4 mm threaded portion = 7 mm total

| PFT3 (1 per box)

| PFT3-5 (5 per box)

actual size

1.5 mm x 4 mm *polished neck*

+ 4 mm threaded portion = 8 mm total

| PFT4 (1 per box)

| PFT4-5 (5 per box)

actual size

1.5 mm x 5 mm *polished neck*

+ 4 mm threaded portion = 9 mm total

| PFT5 (1 per box)

| PFT5-5 (5 per box)

Fully Threaded Self-Drilling Tenting Screws

actual size

1.5 mm x 8 mm

| PFT8 (1 per box)

actual size

1.5 mm x 10 mm

| PFT10 (1 per box)

Pro-Fix™ Bone Fixation

Precision Fixation System

Pro-fix™ Bone Fixation Screws are designed with finer pitched, self-tapping threads that give the screws greater clamping force while using less driver torque. The screws' threads are equipped with a cutting flute that allows for easier insertion into harder bone. The screws are placed into a 1.2 mm pre-drilled pilot hole.

Bone Fixation Kit

| PFBK12S

Autoclavable Tecapro™ storage tray w/ screw organizer dial
Stainless steel driver handle
76 mm cruciform driver blade
56 mm cruciform driver blade
1.2 mm diameter latch type pilot drill
(2) 1.5 x 8 mm bone fixation screws
(4) 1.5 x 10 mm bone fixation screws
(4) 1.5 x 12 mm bone fixation screws
(2) 1.5 x 14 mm bone fixation screws

For individual Pro-Fix™ driver and container components, see page 32.

Self-Tapping Bone Fixation Screws

1.5 mm x 8 mm

PFB8	(1 per box)
PFB8-5	(5 per box)

actual size

1.5 mm x 10 mm

PFB10	(1 per box)
PFB10-5	(5 per box)

actual size

1.5 mm x 12 mm

PFB12	(1 per box)
PFB12-5	(5 per box)

actual size

1.5 mm x 14 mm

PFB14	(1 per box)
PFB14-5	(5 per box)

actual size

Swann-Morton®

Premium Micro-Serrated Blades

"The Swann-Morton® blades have several advantages: First they cut, and ***they cut clean and easy***. Secondly, ***their shape is perfect***. The 15c is like a microsurgical blade, cutting precisely with its spiky tip. The 15 blade has a long, perfectly angulated blade that can be used very safely for eliminating periosteal bundles around the nerve. I use the 15 blade for this and for cutting through the periosteum on the third zone of the lingual flap."

Istvan Urban, DMD, MD, PhD
Periodontist
Oral and Maxillofacial Surgeon

SWANN-MORTON® BLADE EDGE DESIGN



Unique cutting-edge design delivers a consistently sharp blade.

COMPETITOR BLADE EDGE DESIGN

While initially sharp, this edge can deteriorate faster.

15 Blade

01SM15	Stainless Steel	(100 per box)
00SM15	Carbon Steel	(100 per box)

15C Blade

01SM15C	Stainless Steel	(100 per box)
00SM15C	Carbon Steel	(100 per box)

12D Blade

01SM12D	Stainless Steel	(100 per box)
00SM12D	Carbon Steel	(100 per box)

SMOOTH RAZOR EDGE SUPPORTED BY A MICRO-SERRATED EDGE

Maintains a consistently sharp blade

EDGE DESIGN DELIVERS A TACTILE SENSITIVITY

Improves depth control while providing equal, smooth tissue margins

Meisinger since 1888

This system is used for the fixation of membranes in order to avoid micro-mobility of the graft. Mini-threads on the pins make them a screw/pin hybrid, and the sharp tip allows precise placement into cortical bone. The threads increase the surface area of the shaft, resulting in pin stability, making removal possible with the included screwdriver.

Master-Pin-Control (34) **Pins** | BMP00
Master-Pin-Control Basic (10) **Pins** | BMPBA

- Master-Pin-Tray
- Screw Driver For Pin Removal
- Fixation Holder
- Initial Bur
- Twist Drills: (2) 0.6 mm twist drills, (2) 0.8 mm twist drills

Master-Pin Longer Screwdriver

Designed to make pin removal easier in hard to reach areas

MP15	(1 per box)	

Membrane Fixation Pin Holder

MP12	(1 per box)	

Replacement Pins

MP10	(10 per box)	
MP50	(50 per box)	
MP100	(100 per box)	

Decortication Bur

1.2 mm diameter x 4.0 mm long decortication bur with drill stop

203S-012-RA	(2 per box)	

Master-Core Professional (20) **Trephines** | BMCPR
Master-Core-Basic (10) **Trephines** | BMCBA

Meisinger

The Micro Screw System is used for the minimally invasive and safe fixation of cortical bone grafts. The osteosynthesis screws, developed with Prof. Dr. Fouad Khoury, are made from a special surgical stainless steel and, with their diameters of 1.0 mm and 1.2 mm, they ensure that bone plates are fixed in a stable and secure manner. The Micro Screw System can be used for vertical, horizontal, and three-dimensional bone augmentation with different defect sizes.

Micro Screw System (24) Screws | BMS00

Micro Screw System Basic (18) Screws | BMSBA

Micro Screws

1 x 4 mm	MSS10-040	(3 per box)
1 x 6 mm	MSS10-060	(3 per box)
1 x 8 mm	MSS10-080	(3 per box)
1 x 10 mm	MSS10-100	(3 per box)
1.2 x 4 mm	MSS12-040	(3 per box)
1.2 x 6 mm	MSS12-060	(3 per box)
1.2 x 8 mm	MSS12-080	(3 per box)
1.2 x 10 mm	MSS12-100	(3 per box)
1 x 12 mm	MSS10-120	(3 per box)
1 x 14 mm	MSS10-140	(3 per box)
1.2 x 12 mm	MSS12-120	(3 per box)
1.2 x 14 mm	MSS12-140	(3 per box)

Dr. Fouad Khoury Trephine Ejection Kit | BTE00

The Trephine Ejection Kit, developed with Prof. Dr. Fouad Khoury, offers efficiency, flexibility, and precision for the extraction of bone cylinders. The two-part internally cooled trephines allow for safe, minimally invasive extraction of local bone cylinders. The inner cooling helps prevent overheating damage to bone tissue, while the removable trephine working parts allow for an easy sampling of bone cylinders, which can be pushed out with the aid of the ejection needle or precision-fitted ejection instruments.

Master-Mill | BKM00

NEW

Versah®
THE OSSEODENSIFICATION COMPANY

VT5® Series

Regular Densah® Burs

VT1525	1.5 - 2.5 mm Densah® Bur G3
VT2535	2.5 - 3.5 mm Densah® Bur G3
VT3545	3.5 - 4.5 mm Densah® Bur G3
VT4555	4.5 - 5.5 mm Densah® Bur G3
VT5565	5.5 - 6.5 mm Densah® Bur G3

Short Densah® Burs

VT1525-S	1.5 - 2.5 mm Densah® Bur G3 33mm Length
VT2535-S	2.5 - 3.5 mm Densah® Bur G3 33mm Length
VT3545-S	3.5 - 4.5 mm Densah® Bur G3 33mm Length
VT4555-S	4.5 - 5.5 mm Densah® Bur G3 33mm Length

VT8® Series

Regular Densah® Burs

VT1828	1.8 - 2.8 mm Densah® Bur G3
VT2838	2.8 - 3.8 mm Densah® Bur G3
VT3848	3.8 - 4.8 mm Densah® Bur G3
VT4858	4.8 - 5.8 mm Densah® Bur G3
VT5868	5.8 - 6.8 mm Densah® Bur G3

Short Densah® Burs

VT1828-S	1.8 - 2.8 mm Densah® Bur G3 33mm Length
VT2838-S	2.8 - 3.8 mm Densah® Bur G3 33mm Length
VT3848-S	3.8 - 4.8 mm Densah® Bur G3 33mm Length
VT4858-S	4.8 - 5.8 mm Densah® Bur G3 33mm Length

VS8® Series

Regular Densah® Burs

VS2228	2.2 - 2.8 mm Densah® Bur G3
VS3238	3.2 - 3.8 mm Densah® Bur G3
VS4248	4.2 - 4.8 mm Densah® Bur G3
VS5258	5.2 - 5.8 mm Densah® Bur G3

Densah® Bur Technology - Pilot Drills

VPLTT	1.6 mm Tapered Pilot Drill G3
VPLTT-S	1.6 mm Tapered Pilot Drill G3 33 mm length

The Innovation That Makes Osseodensification Possible

The Densah® Bur uses advanced, patented technology designed to preserve and densify bone. This technique, known as Osseodensification, autografts bone along the entire length of the osteotomy through a hydrodynamic process which can help surgeons facilitate multiple clinical procedures such as Sinus Lift, Ridge, and Molar Septum expansion.

NEW

Densah® Bur Technology - Accessories

|	VXTR	Versah® Extender
|	VPLP	Versah® Parallel Pin 1.5 mm - 3.5 mm
|	VPLP-XL	Versah® Parallel Pin 3.5 mm - 5.5 mm

Total Universal Densah® Bur Kit | TUDBK

DLX Holder, Regular & Short Pilot Drills, Regular & Short Densah® Burs, (3) Parallel Pins, (3) XL Parallel Pins, Drill Extender

Universal Densah® Bur Kit X | UDBKX

X Holder, Regular Pilot Drill, Regular Densah® Burs, (2) Parallel Pins, (2) XL Parallel Pins, Drill Extender

Universal Regular Densah® Bur Kit | URDBK

DLX Holder, Regular Pilot Drill, Regular Densah® Burs, (3) Parallel Pins, (3) XL Parallel Pins, Drill Extender

Universal Short Densah® Bur Kit | USDBK

DLX Holder, Short Pilot Drill, Short Densah® Burs, (3) Parallel Pins, (3) XL Parallel Pins, Drill Extender

Densah® Sinus Lift Kit | DSLK

Sinus Kit Holder, Guided Pilot Drill, Guided Densah®Burs, Guided Short Densah® Burs, Large Vertical Stops® (3mm, 5mm, 8mm, 10mm, 11.5mm)

VT5® Series - Guided

Guided Densah® Burs

VT1525-G	1.5 - 2.5 mm Guided Densah® Bur G4
VT2535-G	2.5 - 3.5 mm Guided Densah® Bur G4
VT3545-G	3.5 - 4.5 mm Guided Densah® Bur G4
VT4555-G	4.5 - 5.5 mm Guided Densah® Bur G4

Guided Short Densah® Burs

VT1525-GS	1.5 - 2.5 mm Guided Densah® Bur G4 35 mm Length
VT2535-GS	2.5 - 3.5 mm Guided Densah® Bur G4 35 mm Length
VT3545-GS	3.5 - 4.5 mm Guided Densah® Bur G4 35 mm Length
VT4555-GS	4.5 - 5.5 mm Guided Densah® Bur G4 35 mm Length

VT8® Series - Guided

Guided Densah® Burs

VT1828-G	1.8 - 2.8 mm Densah® Bur G4
VT2838-G	2.8 - 3.8 mm Densah® Bur G4
VT3848-G	3.8 - 4.8 mm Densah® Bur G4
VT4858-G	4.8 - 5.8 mm Densah® Bur G4

Guided Short Densah® Burs

VT1828-GS	1.8 - 2.8 mm Densah® Bur G4 35 mm Length
VT2838-GS	2.8 - 3.8 mm Densah® Bur G4 35 mm Length
VT3848-GS	3.8 - 4.8 mm Densah® Bur G4 35 mm Length
VT4858-GS	4.8 - 5.8 mm Densah® Bur G4 35 mm Length

VS8® Series - Guided

Guided Densah® Burs

VS2228-G	2.2 - 2.8 mm Densah® Bur G4
VS3238-G	3.2 - 3.8 mm Densah® Bur G4
VS4248-G	4.2 - 4.8 mm Densah® Bur G4
VS5258-G	5.2 - 5.8 mm Densah® Bur G4

Guided Densah® Bur Technology - Pilot Drills

VPLTT-G	1.6 mm Tapered Pilot Drill G4
VPLTT-GS	1.6 mm Short Guided Tapered Pilot Drill

Small Vertical Key

VTSP-S-O-3	3 mm Drilling Depth
VTSP-S-O-5	5 mm Drilling Depth
VTSP-S-O-8	8 mm Drilling Depth
VTSP-S-O-10	10 mm Drilling Depth
VTSP-S-O-11.5	11.5 mm Drilling Depth
VTSP-S-O-13	13 mm Drilling Depth
VTSP-S-O-15	15 mm Drilling Depth

Medium Vertical Key

VTSP-M-O-3	3 mm Drilling Depth
VTSP-M-O-5	5 mm Drilling Depth
VTSP-M-O-8	8 mm Drilling Depth
VTSP-M-O-10	10 mm Drilling Depth
VTSP-M-O-11.5	11.5 mm Drilling Depth
VTSP-M-O-13	13 mm Drilling Depth
VTSP-M-O-15	15 mm Drilling Depth

Large Vertical Key

VTSP-L-O-3	3 mm Drilling Depth
VTSP-L-O-5	5 mm Drilling Depth
VTSP-L-O-8	8 mm Drilling Depth
VTSP-L-O-10	10 mm Drilling Depth
VTSP-L-O-11.5	11.5 mm Drilling Depth
VTSP-L-O-13	13 mm Drilling Depth
VTSP-L-O-15	15 mm Drilling Depth

Extra-Large Vertical Key

VTSP-XL-O-3	3 mm Drilling Depth
VTSP-XL-O-5	5 mm Drilling Depth
VTSP-XL-O-8	8 mm Drilling Depth
VTSP-XL-O-10	10 mm Drilling Depth
VTSP-XL-O-11.5	11.5 mm Drilling Depth
VTSP-XL-O-13	13 mm Drilling Depth
VTSP-XL-O-15	15 mm Drilling Depth

C-Guides

VCG-S	Small, Up to 3.25 mm Implant
VCG-M	Medium, Up to 4.3 mm Implant
VCG-L	Large, Up to 5.3 mm Implant
VCG-XL	XL, Up to 6.2 mm Implant

NEW

Universal Guided Surgery Densah® Bur Kit

| UGSDBK

G-Stop® Holder, Guided Pilot Drill, Guided Short Pilot Drill, Guided Densah® Burs, Guided Short Densah® Burs, (3) Parallel Pins, (3) XL Parallel Pins, Drill Extender, Vertical G-Stops® depths: 3mm, 5mm, 8mm, 10mm, 11.5mm, 13mm, 15mm (S, M, L, XL)

Universal ZGO™ Densah® Bur Kit

| UZDBK

ZGO™ Holder, ZGO™ Tapered Pilot Drills (65 mm & 90 mm), ZGO™ Densah®Burs (8)

ZGO™ Densah® Bur Technology - Pilot Drills

ZPLTT-65	Zygoma Tapered Pilot Drill 65 mm
ZPLTT-90	Zygoma Tapered Pilot Drill 90 mm

ZGO™ Densah® Burs

ZGO™ Densah® Burs 65 mm

ZT1525-65	1.5 - 2.5 mm ZGO™ Densah® Bur 65mm Length
ZT2030-65	2.0 - 3.0 mm ZGO™ Densah® Bur 65mm Length
ZT2535-65	2.5 - 3.5 mm ZGO™ Densah® Bur 65mm Length
ZT3040-65	3.0 - 4.0 mm ZGO™ Densah® Bur 65mm Length

ZGO™ Densah® Burs 90 mm

ZT1525-90	1.5 - 2.5 mm ZGO™ Densah® Bur 90mm Length
ZT2030-90	2.0 - 3.0 mm ZGO™ Densah® Bur 90mm Length
ZT2535-90	2.5 - 3.5 mm ZGO™ Densah® Bur 90mm Length
ZT3040-90	3.0 - 4.0 mm ZGO™ Densah® Bur 90mm Length

Micross & Micross Curve

Minimally invasive cortical bone collectors

Holds up to 0.25 cc at a time not actual size

Micross	|	**4049**	(1 sterile scraper per package)

The cannula's 5 mm external diameter allows the Micross to be easily inserted into tissue tunnels.

APPLICATIONS

- Extraction defects
- Periodontal defects
- Sinus lift procedures

HARVESTING SITES

- Oblique external line with tunnel
- Lingual bone
- Sinus window
- Palate
- Zygomatic area with tunnel
- Small areas near the defect

Holds up to 0.25 cc at a time not actual size

Micross Curve	|	**5489**	(1 sterile scraper per package)

The Micross Curve's ergonomical 27º design allows the user to easily reach retromandibular, palatal, and other challenging areas.

5.0 mm DIAMETER

Device can be used in hard-to-reach intraoral sites using a tunnel technique

BONE IS COLLECTED WITH COAGULATED BLOOD

Graft has high biological plasticity, making it easy to handle and mold

SUPERIOR HARVESTING METHOD

The manual harvesting technique allows graft to retain cell viability that can be compromised with other harvesting techniques that mill, grind, or potentially overheat bone

SAFE

The disposable scraper is sterile and allows clinicians to harvest autogenous bone, which eliminates any chance of disease transmission

Safescraper® Twist - Curve

Versatile cortical bone collector

A 160º blade allows clinicians to collect bone from any bony surface.

The Safescraper® Twist's transparent chamber holds up to 2.5 cc of bone that can be used alone or mixed in combination with other graft materials.

not actual size

Holds up to 2.5 cc at a time

| **3987** (3 sterile scrapers per package)

APPLICATIONS

- Extraction defects
- Periodontal defects
- Sinus lift procedures
- Ridge augmentation

HARVESTING SITES

- Oblique external line with tunnel
- Ramus
- Mandibular symphysis
- Sinus window
- Lingual bone
- Zygomatic area
- Nasal spine
- Palate
- Small areas near the defect

Safescraper® Twist - Curve Volumizer

Versatile cortical bone collector

Double grooves on the blade allow for three collecting surfaces.

The Volumizer offers the same performance while collecting a larger volume of autologous cortical bone.

not actual size

Holds up to 2.5 cc at a time

| **5501** (3 sterile scrapers per package)

Selection of Applicable References

MEMBRANES

• Antonious M, Couso-Queiruga E, Barwacz C, Gonzalez-Martin O, Avila-Ortiz G. Evaluation of a Minimally Invasive Alveolar Ridge Reconstruction Approach in Postextraction Dehiscence Defects: A Case Series. Int J Periodontics Restorative Dent. 2021 May-Jun;41(3):335-345.

• Al Hugail AM, Mealey BL, Walker C, Al Harthi S, Duong M, Noujeim M, Lasho DJ, Prihoda TJ, Huynh-Ba G. Evaluation of healing at molar extraction sites with ridge preservation using a non-resorbable dense polytetrafluoroethylene membrane: A four-arm cohort prospective study. Clin Exp Dent Res. 2021 Jun 6. Epub ahead of print.

• Maiorana MC, Fontana F, Dal Polo MR, Pieroni S, Ferrantino L, Poli PP, Simion M. Dense Polytetrafluoroethylene Membrane versus Titanium Mesh in Vertical Ridge Augmentation: Clinical and Histological Results of a Split-mouth Prospective Study. J Contemp Dent Pract 2021; 22 (5):465-472.

• Urban IA, Barootchi S, Tavelli L, Wang HL. Inter-Implant Papilla Reconstruction via a Bone and Soft Tissue Augmentation: A Case Report with a Long-Term Follow-up. Int J Periodontics Restorative Dent. 2021 Nov-Dec;41(2):169-175.

• Amaral Valladao CA JR, Freitas Monteiro M, Joly JC. Guided bone regeneration in staged vertical and horizontal bone augmentation using platelet-rich fibrin associated with bone grafts: a retrospective clinical study. Int J Implant Dent. 2020 Oct 17;6(1):72.

• Nelson AC, Mealey BL. A randomized controlled trial on the impact of healing time on wound healing following ridge preservation using a 70%/30% combination of mineralized and demineralized freeze-dried bone allograft. J Periodontol. 2020 Oct;91(10):1256-1263.

• Pistilli R, Simion M, Barausse C, Gasparro R, Pistilli V, Bellini P, Felice P. Guided Bone Regeneration with Nonresorbable Membranes in the Rehabilitation of Partially Edentulous Atrophic Arches: A Retrospective Study on 122 Implants with a 3- to 7-Year Follow-up. Int J Periodontics Restorative Dent. Sep/Oct 2020;40(5):685-692.

• Windisch P, Orban K, Salvi Ge, Sculean A, Molnar B. Vertical-guided bone regeneration with a titanium-reinforced d-PTFE membrane utilizing a novel split-thickness flap design: a prospective case series. Clin Oral Investig. 2020 Oct 10. Epub ahead of print.

• Cucchi A, Vignudelli E, Fiorino A, Pellegrino G, Corinaldesi G. Vertical ridge augmentation (VRA) with Ti-reinforced d-PTFE membranes or Ti meshes and collagen membranes: 1-year results of a randomized clinical trial. Clin Oral Implants Res. 2020 Oct 5. Epub ahead of print.

• Avila-Ortiz G, Gubler M, Romero-Bustillos M, Nicholas CL, Zimmerman MB, Barwacz CA. Efficacy of Alveolar Ridge Preservation: A Randomized Controlled Trial. J Dent Res. 2020 Feb 12:22034520905660.

• Wen SC, Barootchi S, Huang WX, Wang HL. Time analysis of alveolar ridge preservation using a combination of mineralized bone-plug and dense-polytetrafluoroethylene membrane: A histomorphometric study. J Periodontol. 2020 Feb;91(2):215-222.

• Ibraheem AG, Blanchard SB. Alveolar Ridge Augmentation Around Exposed Mandibular Dental Implant with Histomorphometric Analysis. Clin Adv Periodontics. 2020 Jan 22.

• Koidou VP, Chatzopoulos GS, Johnson D. The "Combo Technique": A Case Series Introducing the Use of a d-PTFE Membrane in Immediate Postextraction Guided Bone Regeneration. J Oral Implantol. 2019 Dec;45(6):486-493.

• Wen SC, Huang WX, Wang HL. Regeneration of Peri-implantitis Infrabony Defects: Report on Three Cases. Int J Periodontics Restorative Dent. 2019 Sep/Oct;39(5):615-621.

• Mazor Z, Horowitz RA, Prasad H, Kotsakis GA. Healing Dynamics Following Alveolar Ridge Preservation with Autologous Tooth Structure. Int J Periodontics Restorative Dent. 2019 Sep/Oct;39(5):697-702.

• Nguyen V, Von Krockow N, Pouchet J, Weigl PM. Periosteal Inhibition Technique for Alveolar Ridge Preservation as It Applies to Implant Therapy. Int J Periodontics Restorative Dent. 2019 Sep/Oct;39(5):737-744.

• Sabe-Alarab M, Al-Essa H, Jaber F, Shomal Y, Kharfan J. Alveolar ridge preservation with d-ptfe membrane a randomized controlled trial. Int J Recent Sci Res. 10(09), pp.34658-34664.

• Cucchi A, Sartori M, Aldini NN, Vignudelli E, Corinaldesi G. A Proposal of Pseudo-periosteum Classification After GBR by Means of Titanium-Reinforced d-PTFE Membranes or Titanium Meshes Plus Cross-Linked Collagen Membranes. Int J Periodontics Restorative Dent. 2019 Jul/Aug;39(4):e157-e165.

• Faciola Pessôa De Oliveira PG, Pedroso Bergamo ET, Bordin D, Arbex L, Konrad D, Gil LF, Neiva R, Tovar N, Witek L, Coelho PG. Ridge Architecture Preservation Following Minimally Traumatic Exodontia Techniques and Guided Tissue Regeneration. Implant Dent. 2019 Aug;28(4):319-328.

• Gallo P, Díaz-Báez D. Management Of 80 Complications In Vertical And Horizontal Ridge Augmentation With Nonresorbable Membrane (d-PTFE): A Cross-Sectional Study. Int J Oral Maxillofac Implants. 2019 July/August;34(4):927–935.

• Urban I, Montero E, Monje A, Sanz-Sanchez I. Effectiveness of vertical ridge augmentation interventions: A systematic review and meta-analysis. J Clin Periodontol. 2019 Jun;46 Suppl 21:319-339.

• Cheng A, Berridge J, McGary R, Erley K, Johnson T. The Extraction Socket Management Continuum: A Hierarchical Approach to Dental Implant Site Development. Clinical Advances in Periodontics, Vol. 9, No. 2, June 2019.

• Urban I, Montero E, Monje A, Sanz-Sánchez I. Effectiveness of vertical ridge augmentation interventions: A systematic review and meta-analysis. J Clin Periodontol. 2019 Jun;46 Suppl 21:319-339.

• Phillips DJ, Swenson DT, Johnson TM. Buccal bone thickness adjacent to virtual dental implants following guided bone regeneration. J Periodontol. 2019 Jun;90(6):595-607.

• Altiparmak N, Akdeniz SS. Primary closure versus open membrane technique in augmentation of deficient alveolar ridges Int J Oral Maxillofac Surg., Vol. 48, Supplement 1, 43, May 01, 2019.

• Mertens C, Braun S, Krisam J, Hoffmann J. The influence of wound closure on graft stability: An in vitro comparison of different bone grafting techniques for the treatment of one-wall horizontal bone defects. Clin Implant Dent Relat Res. 2019 Apr;21(2):284-291.

• Wu IH, Bakhshalian N, Galaustian R, Naini RB, Min S, Freire M, Zadeh HH. Retrospective Analysis of the Outcome of Ridge Preservation with Anorganic Bovine Bone Mineral: Marginal Bone Level at Implants Placed Following Healing of Grafted Extraction Sockets. Int J Periodontics Restorative Dent. 2019 Jan/Feb;39(1):131-140.

• Phillips DJ, Swenson DT, Johnson TM. Buccal bone thickness adjacent to virtual dental implants following guided bone regeneration. J Periodontol. 2018 Dec 21.. [Epub ahead of print]

• Mendoza-Azpur G, Gallo P, Mayta-Tovalino F, Alva R, Valdivia E. A Case Series of Vertical Ridge

Selection of Applicable References

Augmentation Using a Nonresorbable Membrane: A Multicenter Study. Int J Periodontics Restorative Dent. 2018 Nov/Dec;38(6):811-816.

• Sun DJ, Lim HC, Lee DW. Alveolar ridge preservation using an open membrane approach for sockets with bone deficiency: A randomized controlled clinical trial. Clin Implant Dent Relat Res. 2018 Nov 5.[Epub ahead of print]

• Changi KK, Greenstein G.Cytocone Procedure: Conservative Repair of a Buccal Plate Dehiscence in Preparation for Implant Placement. Compend Contin Educ Dent. 2018 May;39(5):294-299.

• Plonka AB, Urban IA, Wang HL. Decision Tree for Vertical Ridge Augmentation. Int J Periodontics Restorative Dent. 2018 Mar/Apr;38(2):269-275.

• Johnson M, Baron D. Tunnel Access for Guided Bone Regeneration in the Maxillary Anterior Clinical Advances in Periodontics. Vol 8. No 1. March 2018.

• Bakhshalian N, Freire M, Min S, Wu I, Zadeh HH. Retrospective Analysis of the Outcome of Ridge Preservation with Anorganic Bovine Bone Minerals: Microcomputed Tomographic Assessment of Wound Healing in Grafted Extraction Sockets. Int J Periodontics Restorative Dent. 2018 Jan/Feb;38(1):103-111.

• Urban I, Traxler H, Romero-Bustillos M, Farkasdi S, Bartee B, Baksa G, Avila-Ortiz G. Effectiveness of Two Different Lingual Flap Advancing Techniques for Vertical Bone Augmentation in the Posterior Mandible: A Comparative, Split-Mouth Cadaver Study. Int J Periodontics Restorative Dent. 2018 Jan/Feb;38(1):35-40.

• Johnson TM, Berridge JP, Baron D. Protocol for Maintaining Alveolar Ridge Volume in Molar Immediate Implant Sites. Clinical Advances in Periodontics. November 2017, Vol. 7, No. 4, Pages 207-214

• Pistilli R, Checchi V, Sammartino G, Simion M, Felice P. Safe New Approach to the Lingual Flap Management in Mandibular Augmentation Procedures: The Digitoclastic Technique. Implant Dent. 2017 Oct;26(5):790-795.

• Urban IA, Monje A, Lozada J, Wang HL. Principles for Vertical Ridge Augmentation in the Atrophic Posterior Mandible: A Technical Review. Int J Periodontics Restorative Dent. 2017 Sep/Oct;37(5):639-645.

• Cucchi A, Vignudelli E, Napolitano A, Marchetti C, Corinaldesi G. Evaluation of complication rates and vertical bone gain after guided bone regeneration with non-resorbable membranes versus titanium meshes and resorbable membranes. A randomized clinical trial. Clin Implant Dent Relat Res. 2017 Jul 26. Epub ahead of print.

• Gultekin BA, Cansiz E, Borahan MO. Clinical and 3-Dimensional Radiographic Evaluation of Autogenous Iliac Block Bone Grafting and Guided Bone Regeneration in Patients With Atrophic Maxilla. J Oral Maxillofac Surg. 2017 Apr;75(4):709-722.

• Ghensi P, Stablum W, Bettio E, Soldini MC, Tripi TR, Soldini C. Management of the exposure of a dense PTFE (d-PTFE) membrane in guided bone regeneration (GBR): a case report. Oral Implantol (Rome). 2017 Nov 30;10(3):335-342.

• Laurito D, Lollobrigida M, Gianno F, Bosco S, Lamazza L, De Biase A. Alveolar Ridge Preservation with nc-HA and d-PTFE Membrane: A Clinical, Histologic, and Histomorphometric Study. Int J Periodontics Restorative Dent. 2017 Mar/Apr;37(2):283-290.

• Walker CJ, Prihoda TJ, Mealey BL, Lasho DJ, Noujeim M, Huynh-Ba G. Evaluation of healing at molar extraction sites with and without ridge preservation: a randomized controlled clinical trial. J Periodontol. 2017 Mar;88(3):241-249.

• Laurito D, Cugnetto R, Lollobrigida M, Guerra F, Vestri A, Gianno F, Bosco S, Lamazza L, De Biase A. Socket Preservation with d-PTFE Membrane: Histologic Analysis of the Newly Formed Matrix at Membrane Removal. Int J Periodontics Restorative Dent. 2016 Nov/Dec;36(6):877-883.

• Ronda M, Stacchi C. A Novel Approach for the Coronal Advancement of the Buccal Flap. Int J Periodontics Restorative Dent. 2015 Nov-Dec;35(6):795-801.

• Urban IA, Monje A, Wang HL. Vertical Ridge Augmentation and Soft Tissue Reconstruction of the Anterior Atrophic Maxillae: A Case Series. Int J Periodontics Restorative Dent. 2015 SepOct;35(5):613-23.

• Al-Hezaimi K, Iezzi G, Rudek I, Al-Daafas A, Al-Hamdan K, Al-Rasheed A, Javed F, Piattelli A, Wang HL. Histomorphometric Analysis of Bone Regeneration Using a Dual Layer of Membranes (dPTFE Placed Over Collagen) in Fresh Extraction Sites: A Canine Model. J Oral Implantol. 2015 Apr;41(2):188-95.

• Borg TD, Mealey BL. Histologic healing following tooth extraction with ridge preservation using mineralized versus combined mineralized-demineralized freeze-dried bone allograft: a randomized controlled clinical trial. J Periodontol. 2015 Mar;86(3):348-55.

• Cucchi A, Ghensi P. Vertical Guided Bone Regeneration using Titanium-reinforced d-PTFE Membrane and Prehydrated Corticocancellous Bone Graft. Open Dent J. 2014 Nov 14;8:194-200.

• Ronda M, Rebaudi A, Torelli L, Stacchi C. Expanded vs. dense polytetrafluoroethylene membranes in vertical ridge augmentation around dental implants: a prospective randomized controlled clinical trial. Clin Oral Implants Res. 2014 Jul;25(7):859-66.

• Barboza EP, Stutz B, Mandarino D, Rodrigues DM, Ferreira VF. Evaluation of a dense polytetrafluoroethylene membrane to increase keratinized tissue: a randomized controlled clinical trial. Implant Dent. 2014 Jun;23(3):289-94.

• Urban IA, Lozada JL, Jovanovic SA, Nagursky H, Nagy K. Vertical Ridge Augmentation with Titanium-Reinforced, Dense-PTFE Membranes and a Combination of Particulated Autogenous Bone and Anorganic Bovine Bone-Derived Mineral: A Prospective Case Series in 19 Patients. Int J Oral Maxillofac Implants. 2014 Jan-Feb;29(1):185-93.

• Carbonell JM, Martin IS, Santos A, Pujol A, SanzMoliner JD, Nart J. High-density polytetrafluoroethylene membranes in guided bone and tissue regeneration procedures: a literature review. Int J Oral Maxillofac Surg. 2014 Jan;43(1):75-84.

• Vittorini Orgeas G, Clementini M, De Risi V, de Sanctis M. Surgical techniques for alveolar socket preservation: a systematic review. Int J Oral Maxillofac Implants. 2013 Jul-Aug;28(4):1049-61.

• Al-Hezaimi K, Rudek I, Al-Hamdan KS, Javed F, Nooh N, Wang HL. Efficacy of using a dual layer of membrane (dPTFE placed over collagen) for ridge preservation in fresh extraction sites: a micro-computed tomographic study in dogs. Clin Oral Implants Res. Clin Oral Implants Res. 2013 Oct;24(10):1152-7.

• Waasdorp, J, Feldman, S. Bone regeneration around immediate implants utilizing a dense polytetrafluoroethylene membrane without primary closure: A report of 3 cases. J Oral Implantol. 2013;39:355-361.

• Annibali S, Bignozzi I, Sammartino G, La Monaca G, Cristalli MP. Horizontal and Vertical Ridge Augmentation in Localized Alveolar Deficient Sites: A Retrospective Case Series. Implant Dent. 2012 Jun;21(3):175-185.

• Levin B. Immediate temporization of immediate implants in the esthetic zone: Evaluating survival and bone maintenance. Compendium 2011;32:52-62.

Selection of Applicable References

• Barboza EP, Stutz B, Ferreira VF, Carvalho W. Guided bone regeneration using nonexpanded polytetrafluoroethylene membranes in preparation for dental implant placements – A report of 420 cases. Implant Dent. 2010;19:2-7.

• Zafiropoulos GG, Deli G, Bartee BK, Hoffman O. Single-tooth implant placement and loading in fresh and regenerated extraction sockets. Five-year results: A case series using two different implant designs. J Periodontol. 2010;81:604-615.

• Zafiropoulos GG, Hoffmann O, Kasaj A, Willershausen B, Deli G, Tatakis DN. Mandibular molar root resection versus implant therapy: A retrospective nonrandomized study. J Oral Implantol. 2009;35:52-62.

• Fotek PD, Neiva RF, Wang HL. Comparison of dermal matrix and polytetrafluoroethylene membrane for socket bone augmentation: A clinical and histologic study. J Periodontol. 2009;80:776-785.

• Hoffman O, Bartee BK, Beaumont C, Kasaj A, Deli G, Zafiropoulos GG. Alveolar bone preservation in extraction sockets using non-resorbable dPTFE membranes: A retrospective non-randomized study. J Periodontol. 2008;79:1355-1369.

• Barber HD, Lignelli J, Smith BM, Bartee BK. Using a dense PTFE membrane without primary closure to achieve bone and tissue regeneration. J Oral Maxillofac Surg. 2007;65:748-752.

• Walters SP, Greenwell H, Hill M, Drisko C, Pickman K, Scheetz JP. Comparison of porous and non-porous teflon membranes plus a xenograft in the treatment of vertical osseous defects: A clinical reentry study. J Periodontol. 2003;74:1161-1168.

• Bartee BK. Extraction site reconstruction for alveolar ridge preservation. Part 1: Rationale and material selection. J Oral Implantol. 2001;27:187-193.

• Bartee BK. Extraction site reconstruction for alveolar ridge preservation. Part 2: Membrane-assisted surgical technique. J Oral Implantol. 2001;27:194-197.

• Lamb JW III, Greenwell H, Drisko C, Henderson RD, Scheetz JP, Rebitski G. A comparison of porous and non-porous teflon membranes plus demineralized freeze-dried bone allograft in the treatment of class II buccal/lingual furcation defects: A clinical reentry study. J Periodontol. 2001;72:1580-1587.

• Bartee BK. Evaluation of a new polytetrafluoroethylene guided tissue regeneration membrane in healing extraction sites. Compend Contin Educ Dent 1998;19:1256-1264.

• Bartee BK, Carr JA. Evaluation of a high-density polytetrafluoroethylene (n-PTFE) membrane as a barrier material to facilitate guided bone regeneration in the rat mandible. J Oral Implantol. 1995;21:88-95.

• Bartee BK. The use of high-density polytetrafluoroethylene membrane to treat osseous defects: Clinical reports. Implant Dent. 1995;4:21-26.

ZMATRIX™

• Cadenas-Vacas G, Martínez-Rodríguez N, Barona-Dorado C, Sánchez-Labrador L, Cortés-Bretón Brinkmann J, Meniz-García C, et al. Calcium phosphate modified with silicon vs. Bovine hydroxyapatite for alveolar ridge preservation: densitometric evaluation, morphological changes and histomorphometric study. Materials (Basel) [Internet]. 2021;14(4): 940.

• Bruyckere T de, Cosyn J, Younes F, Hellyn J, Bekx J, Cleymaet R, et al. A randomized controlled study comparing guided bone regeneration with connective tissue graft to reestablish buccal convexity: One-year aesthetic and patient-reported outcomes. Clin Oral Implants Res. 2020;31(6): 507-16.

• Redemagni M, Mascetti T, Garlini G. Post-extractive immediate implant placement and immediate provisionalization at sites requiring buccal bone regeneration:. EC Dental Science. 2019(18.6): 1207-16.

• Raz P, Brosh T, Ronen G, Tal H. Tensile properties of three selected collagen membranes. Biomed Res Int. 2019 Dec 5;2019:5163603.

• Sanz-Sanchez I, Wessing B, Polizzi G, et al. Randomized clinical trial comparing two resorbable collagen membranes demonstrates good bone formation and soft tissue healing with GBR at single implant sites with dehiscence defects. J Clin Periodontol 2018;45(S19):19–20 [oral presentation].

• Omar O, Dahlin A, Gasser A, et al. Tissue dynamics and regenerative outcome in two resorbable noncross- linked collagen membranes for guided bone regeneration: A preclinical molecular and histological study in vivo. Clin Oral Impl Res; 2018;29(1):7–19.

• Wessing B, Urban I, Montero E, et al. A multicenter randomized controlled clinical trial using a new resorbable non-cross- linked collagen membrane for guided bone regeneration at dehisced single implant sites: interim results of a bone augmentation procedure. Clin Oral Impl Res; 2017;28(11):e218–e226.

• Jiménez Garcia J, Berghezan S, Caramês JMM, Dard MM, Marques DNS. Effect of cross-linked vs non-cross-linked collagen membranes on bone: A systematic review, J Periodontol. Res. 2017;1–10.

• Wessing B, Emmerich M, Bozkurt A. Horizontal ridge augmentation with a novel resorbable collagen membrane: a retrospective analysis of 36 consecutive patients. Int J Periodontics Restorative Dent 2016;36(2):179–187.

• Gasser A, Wessing B, Eummelen L, et al. Mechanical stability of collagen membranes: an in vitro study. J Dent Res 2016;95(Spec Iss A): 1683.

• Bozkurt A, Apel C, Sellhaus B, et al. Differences in degradation behavior of two non-cross-linked collagen barrier membranes: an in vitro and in vivo study. Clin Oral Impl Res; 2014; 25(12):1403-1411.

• Jäger M, Degistirici O, Knipper A, Fischer J, Sager M, Krauspe R. Bone healing and migration of cord blood-derived stem cells into a critical size femoral defect after xenotransplantation. J Bone Miner Res. 2007;22(8): 1224-33.

RIDGE AUGMENTATION MESH

• Urban IA, Saleh MHA, Ravida A, Forster A, Wang HL, Barath Z. Vertical bone augmentation utilizing a titanium-reinforced PTFE mesh: A multi-variate analysis of influencing factors. Clin Oral Implants Res. 2021 Mar 31. Epub ahead of print.

• Urban I, Montero E, Monje A, Sanz-Sanchez I. Effectiveness of vertical ridge augmentation interventions: A systematic review and meta-analysis. J Clin Periodontol. 2019 Jun;46 Suppl 21:319-339.

COMBINATION ALLOGRAFT

• Antonious M, Couso-Queiruga E, Barwacz C, Gonzalez-Martin O, Avila-Ortiz G. Evaluation of a Minimally Invasive Alveolar Ridge Reconstruction Approach in Postextraction Dehiscence Defects: A Case Series. Int J Periodontics Restorative Dent. 2021 May-Jun;41(3):335-345.

• Nelson AC, Mealey BL. A randomized controlled trial on the impact of healing time on wound healing following ridge preservation using a 70%/30% combination of mineralized and demineralized freeze-dried bone allograft. J Periodontol. 2020 Oct;91(10):1256-1263.

• Cucchi A, Sartori M, Aldinia NN, Vignudelli E, Corinaldesi G. A Proposal of Pseudo-periosteum Classification After GBR by Means of Titanium-

Selection of Applicable References

Reinforced d-PTFE Membranes or Titanium Meshes Plus Cross-Linked Collagen Membranes. Int J Periodontics Restorative Dent. 2019 Jul/Aug;39(4):e157-e165.

• Demetter RS, Calahan BG, Mealey BL. Histologic Evaluation of Wound Healing After Ridge Preservation With Cortical, Cancellous, and Combined Cortico-Cancellous Freeze-Dried Bone Allograft: A Randomized Controlled Clinical Trial. J Periodontol. 2017 Sep;88(9):860-868.

• Chan HL, Benavides E, Tsai CY, Wang HL. A Titanium Mesh and Particulate Allograft for Vertical Ridge Augmentation in the Posterior Mandible: A Pilot Study. Int J Periodontics Restorative Dent. 2015 Jul-Aug;35(4):515-22.

• Borg TD, Mealey BL. Histologic healing following tooth extraction with ridge preservation using mineralized versus combined mineralized-demineralized freeze-dried bone allograft: a randomized controlled clinical trial. J Periodontol. 2015 Mar;86(3):348-55.

XENOGRAFT

• Lai VJ, Michalek JE, Liu Q, Mealey BL. Ridge preservation following tooth extraction using bovine xenograft compared with porcine xenograft: A randomized controlled clinical trial. J Periodontol. 2020 Mar;91(3):361-368.

SUTURE

• Taysi AE, Ercal P, Sismanoglu S. Comparison between tensile characteristics of various suture materials with two suture techniques: an in vitro study. Clin Oral Investig. 2021 Apr 14.

• Abellán D, Nart J, Pascual A, Cohen RE, Sanz-Moliner JD. Physical and Mechanical Evaluation of Five Suture Materials on Three Knot Configurations: An in Vitro Study. Polymers. 2016; 8(4):147

• Silverstein LH, Kurtzman GM, Shatz PC. Suturing for optimal soft-tissue management. J Oral Implantol. 2009;35:82-90.

• Silverstein LH. Suturing principles: Preserving needle edges during dental suturing. PPAD. 2005;17:562-564.

BONE SCRAPERS

• Bacci C, Lucchiari N, Valente M, Della Barbera M, Frigo AC, Berengo M. Intra-oral bone harvesting: two methods compared using histological and histomorphometric assessments. Clin Oral Implants Res. 2011 Jun;22(6):600-5.

• Caubet J, Petzold C, Sáez-Torres C, Morey M, Iriarte JI, Sánchez J, Torres JJ, Ramis JM, Monjo M. Sinus graft with safescraper: 5-year results. J Oral Maxillofac Surg. 2011 Feb;69(2):482-90.

• Trombelli L, Farina R, Marzola A, Itro A, Calura G. GBR and autogenous cortical bone particulate by bone scraper for alveolar ridge augmentation: A 2 case report. Int J Oral Maxillofac Implants. 2008;23:111-116.

• Zaffe D, D'Avenia F. A novel bone scraper for intraoral harvesting: A device for filling small bone defects. Clin Oral Implants Res. 2007;18:525-533.

• Trombelli L, Annunziata M, Belardo S, Farina R, Scabbia A, Guida L. Autogenous bone graft in conjunction with enamel matrix derivative in the treatment of deep periodontal intra-osseous defects: A report of 13 consecutively treated patients. J Clin Periodontol. 2006;33:69-75.

NOVABONE®

• Bhandari S, Thomas R, Kumar T, Shah R, Mehta DS. Maxillary Sinus Augmentation Using Hydraulic Pressure by Lateral Approach and Simultaneous Implant Placement: Clinicoradiographic Study. Implant Dent. 2019 Oct;28(5):514-519.

• Bodhare GH, Kolte AP, Kolte RA, Shirke PY. Clinical and radiographic evaluation and comparison of bioactive bone alloplast morsels when used alone and in combination with platelet-rich fibrin in the treatment of periodontal intrabony defects-A randomized controlled trial. J Periodontol. 2019 Jun;90(6):584-594.

• Mahesh L, Venkataraman N, Shukla S, Prasad H, Kotsakis GA. Alveolar ridge preservation with the socket-plug technique utilizing an alloplastic putty bone substitute or a particulate xenograft: a histological pilot study. J Oral Implantol. 2015 Apr;41(2):178-83.

• Kotsakis GA, Mazor Z. A simplified approach to the minimally invasive antral membrane elevation technique utilizing a viscoelastic medium for hydraulic sinus floor elevation. Oral Maxillofac Surg. 2015 Mar;19(1):97-101.

• Ioannou AL, Kotsakis GA, Kumar T, Hinrichs JE, Romanos G. Evaluation of the bone regeneration potential of bioactive glass in implant site development surgeries: a systematic review of the literature. Clin Oral Investig. 2015 Mar;19(2):181-91.

• Babbush CA, Kanawati A. Clinical evaluation of 262 osseointegrated implants placed in sites grafted with calcium phosphosilicate putty: a retrospective study. J Oral Implantol. 2015 Feb;41(1):63-9.

• Kotsakis GA, Joachim FP, Saroff SA, Mahesh L, Prasad H, Rohrer MD. Histomorphometric evaluation of a calcium-phosphosilicate putty bone substitute in extraction sockets. Int J Periodontics Restorative Dent. 2014 Mar-Apr;34(2):233-9.

• Kher U, Mazor Z, Stanitsas P, Kotsakis GA. Implants placed simultaneously with lateral window sinus augmentation using a putty alloplastic bone substitute for increased primary implant stability: a retrospective study. Implant Dent. 2014 Aug;23(4):496-501.

• Kotsakis GA, Salama M, Chrepa V, Hinrichs JE, Gaillard P. A randomized, blinded, controlled clinical study of particulate anorganic bovine bone mineral and calcium phosphosilicate putty bone substitutes for socket preservation. Int J Oral Maxillofac Implants. 2014 Jan-Feb;29(1):141-51.

• Jodia K, Sadhwani BS, Parmar BS, Anchlia S, Sadhwani SB. Sinus elevation with an alloplastic material and simultaneous implant placement: a 1-stage procedure in severely atrophic maxillae. J Maxillofac Oral Surg. 2014 Sep;13(3):271-80.

• Kim DM, Nevins M, Camelo M, Nevins ML, Schupbach P, Rodrigues VS, Fiorellini JP. Human histologic evaluation of the use of the dental putty for bone formation in the maxillary sinus: case series. J Oral Implantol. 2012 Aug;38(4):391-8.

• Lanka M, Salama M, Kurtzman, Gregori.Socket grafting with calcium phosphosilicate alloplast putty: a histomorphometric evaluation. Compend Contin Educ Dent. 2012 Sep;33(8):e109-15.

• Gonshor A, Saroff S, Anderegg C, Joachim F, Charon J, Prasad H, Katta S. Histologic and Clinical Evaluation of a Bioactive Calcium Phosphosilicate Bone Graft Material in Postextraction Alveolar Sockets. Int J Oral Imp and Clin Res. 2011;2(2): 79-84.

• Hench L. The story of Bioglass. J Mater Sci Mater Med. 2006 Nov;17(11):967-78.

• Xynos ID, Edgar AJ, Buttery LD, Hench LL, Polak JM. Gene-expression profiling of human osteoblasts following treatment with the ionic products of Bioglass 45S5 dissolution. J Biomed Mater Res. 2001 May;55(2):151-7.

We're here to help.

888.796.1923

sales@osteogenics.com

Contact us with questions or to place an order...*or shop our products online!*

TOLL-FREE	1.888.796.1923 US & Canada only
INTERNATIONAL	+1 806.796.1923
FAX	806.796.0059
EMAIL	sales@osteogenics.com
WEBSITE	www.osteogenics.com
ADDRESS	Osteogenics Biomedical, Inc. 4620 71st Street | Building 78-79 Lubbock, TX 79424

UNRIVALED
REGEN SOLUTIONS

For 30 years *Osteogenics' mission has been to provide best-in-class, cutting-edge, predictable dental bone grafting products. With ridge augmentation, ridge preservation, sinus elevation, and suturing procedures in mind, our product portfolio has been able to expand and evolve to fit the needs of our loyal customers. Ultimately, developing unrivaled regenerative products means providing* ***a brighter future for our patients.***